

12-30-2013

## Activity of tobramycin and polymyxin-E against *Pseudomonas aeruginosa* biofilm coated medical grade endotracheal tubes.

Keiko Tarquinio  
*University of Rhode Island*

Confreda

James Shurko  
*University of Rhode Island*

Kerry L. LaPlante  
*University of Rhode Island, kerrylaplante@uri.edu*

Follow this and additional works at: [https://digitalcommons.uri.edu/php\\_facpubs](https://digitalcommons.uri.edu/php_facpubs)

---

### Citation/Publisher Attribution

Tarquinio, K., Confreda, K., Shurko, J., & LaPlante, K. (2013). Activity of tobramycin and polymyxin-E against *Pseudomonas aeruginosa* biofilm coated medical grade endotracheal tubes. *Antimicrob. Agents Chemother.*, 58(3), 1723-1729. doi: 10.1128/AAC.01178-13.  
Available at: <http://dx.doi.org/10.1128/AAC.01178-13>

This Article is brought to you by the University of Rhode Island. It has been accepted for inclusion in Pharmacy Practice and Clinical Research Faculty Publications by an authorized administrator of DigitalCommons@URI. For more information, please contact [digitalcommons-group@uri.edu](mailto:digitalcommons-group@uri.edu). For permission to reuse copyrighted content, contact the author directly.

---

**Activity of tobramycin and polymyxin-E against *Pseudomonas aeruginosa* biofilm coated medical grade endotracheal tubes.**

The University of Rhode Island Faculty have made this article openly available.  
Please let us know how Open Access to this research benefits you.

This is a pre-publication author manuscript of the final, published article.

**Terms of Use**

This article is made available under the terms and conditions applicable towards Open Access Policy Articles, as set forth in our [Terms of Use](#).

1 **Activity of tobramycin and polymyxin-E against *Pseudomonas aeruginosa* biofilm**  
2 **coated medical grade endotracheal tubes**

3

4 **Keiko Tarquinio<sup>1,2,3,5</sup>, Kelsey Confreda<sup>1,2</sup>, James Shurko<sup>3</sup>, Kerry LaPlante<sup>2,3,4,#</sup>**

5

6 <sup>1</sup>Pediatric Critical Care Medicine, Hasbro Children's Hospital, Rhode Island Hospital,  
7 Providence, RI

8 <sup>2</sup>Rhode Island Infectious Diseases (RIID) Research Program, Providence Veterans Affairs  
9 Medical Center, Providence, RI

10 <sup>3</sup>University of Rhode Island, Department of Pharmacy Practice, Kingston, RI

11 <sup>4</sup>Department of Medicine, <sup>5</sup>Department of Pediatrics, Warren Alpert Medical School of  
12 Brown University, Providence, RI

13

14 **Running Title:** Tobramycin and polymyxin-E against *P. aeruginosa*

15

16 **#Corresponding Author:** Kerry L. LaPlante, Pharm.D. Associate Professor of Pharmacy,  
17 University of Rhode Island, Veterans Affairs Medical Center (151) ; Research Building 35  
18 ; 830 Chalkstone Avenue; Providence, RI 02908, office: 401-273-7100 x2339; fax: 401-  
19 457-3305; e-mail: KerryLaPlante@uri.edu

20 **ABSTRACT**

21 Indwelling medical devices have become a major source of nosocomial infections;  
22 especially *Pseudomonas aeruginosa* (*P. aeruginosa*) infection, which remain the most  
23 common cause of ventilator associated pneumonia (VAP) in neonates and children. Using  
24 medical grade polyvinyl chloride endotracheal tubes (ETTs), the activity of tobramycin and  
25 polymyxin-E was quantified in a simulated prevention and treatment static time kill model  
26 using biofilm forming *P. aeruginosa*. The model simulated three clinical conditions: 1)  
27 planktonic bacteria in the presence of antibiotics, tobramycin and polymyxin-E, without  
28 ETTs, 2) planktonic bacteria grown in the presence of *P. aeruginosa*, antibiotic and ETTs  
29 (simulating prevention) and 3) a 24h formed *P. aeruginosa* biofilm on ETTs prior to  
30 antibiotic exposure (simulating treatment). In the model simulating “prevention”  
31 (conditions 1 and 2 above), tobramycin alone or in combination with polymyxin-E was  
32 more bactericidal than polymyxin-E alone at 24 hours using a concentration greater than 2  
33 times the minimum inhibitory concentration (MIC). However, after a 24h old biofilm was  
34 allowed to form on the ETTs, neither monotherapy nor combination therapy over 24 hours  
35 exhibited bactericidal or bacteriostatic effects. Against the same pathogens, tobramycin  
36 and polymyxin-E, both alone or in combination exhibited bactericidal activity prior to  
37 biofilm attachment to the ETTs, however no activity was observed once biofilm formed on  
38 ETTs. These findings support surveillance culturing to identify pathogens for a rapid and  
39 targeted approach to therapy, especially when *P. aeruginosa* is a potential pathogen.

## 40 INTRODUCTION

41 Indwelling medical devices are a major source of nosocomial infections. In particular,  
42 patients requiring mechanical ventilation (intubation with an endotracheal tube (ETT)) face  
43 a high probability of contracting one of the most prevalent nosocomial infections,  
44 ventilator associated pneumonia (VAP).(1-3) Neonatal and pediatric populations are at  
45 especially high risk for VAP because the current standard of care involves prolonged  
46 intubation without ETT exchange or tracheostomy, both common practice in adult patients.  
47 In neonates and infants, the inner diameter of the ETT is often 2.5-3.5mm (the size of a thin  
48 straw), which complicates suctioning of secretions and confounds attempts to maintain  
49 patency. Despite aggressive bedside hygiene, *Pseudomonas aeruginosa* (*P. aeruginosa*)  
50 remains one of the most common causes of VAP in intubated children.(2, 4, 5)

51

52 *P. aeruginosa*, often found on indwelling devices such as ETTs, forms a biofilm which  
53 serves as an ideal environment for antibiotic resistance, making VAP difficult to treat.(6, 7)  
54 Biofilm on ETTs is considered to be a reservoir for infecting pathogens derived from  
55 oropharyngeal flora and gastric microaspiration, and is highly correlated with lower airway  
56 infection and subsequent VAP.(8-11) To date, few side-by-side studies have compared  
57 killing activity (defined as 99.9% kill) of tobramycin and polymyxin-E against *P.*  
58 *aeruginosa*, especially in the context of ETT biofilm and VAP.(12-15) The effect of  
59 monotherapy and/or combination therapy (synergistic versus antagonistic activity) must be  
60 assessed when evaluating antimicrobial drug therapy, especially in the presence of medical  
61 grade polyvinyl chloride (PVC) or conventional ETTs. For convenience, most studies  
62 investigating antibiotic susceptibility in formed biofilms have used PVC coupons rather

63 than clinically available medical devices.(16-18) However, most of the coupons made of  
64 PVC are not medical grade and, in many cases, do not contain equivalent plasticizer  
65 content. These differences result in different texture and flexibility between medical grade  
66 PVC products and PVC coupons used in biofilm experiments. Using clinically available  
67 ETTs, this study aimed to both assess the efficacy of antibiotics against planktonic vs.  
68 biofilm formed *P. aeruginosa*, and to identify which antibiotic, alone or combination,  
69 demonstrates the best in vitro activity against *P. aeruginosa* in the context of VAP.

70 **MATERIALS AND METHODS**

71 **Bacterial Isolates.** American Type Culture Collection (ATCC, Manassas, VA, USA) strain  
72 25668 was obtained. Reference strain PAO1 was obtained from Dr. Thomas Murray,  
73 Frank H. Netter MD School of Medicine, Quinnipiac University, North Haven, CT.(19, 20)  
74 Prior to use, all bacteria were stored in tryptic soy broth (TSB; Difco laboratories, Sparks,  
75 MD) with 15% glycerol and frozen at -80 °C. Both strains are prolific biofilm  
76 producers.(21, 22)

77

78 **Antimicrobial Agents.** Commercially available, chemical grade polymyxin-E (lot#  
79 081M1525V) powder and chemical grade tobramycin (lot# 090M1196V) powder was  
80 purchased from Sigma Aldrich (St. Louis, MO). Tobramycin and polymyxin-E powder  
81 were stored at 4°C. Both tobramycin and polymyxin-E were diluted in sterile water and a  
82 fresh stock was made each day, and prior to every experiment. Tobramycin and  
83 polymyxin-E were tested at one, two, four and eight times their respective minimal  
84 inhibitory concentration (MIC) at 0, 4 and 24 hours after inoculation.(23) Cation Adjusted  
85 Mueller-Hinton broth (CA-MHB, Difco Laboratories, Sparks, MD) supplemented with 25  
86 mg/L calcium, 12.5 mg/L magnesium and 0.25% dextrose (Fisher Scientific, Pittsburgh,  
87 PA, USA) was used to obtain a suspension corresponding to 0.7 - 0.8 McFarland standards  
88 to produce an initial starting inocula of  $5.5-6.0 \times 10^6$  colony forming units per milliliter  
89 (CFU/mL). Colony counts were determined using tryptic soy agar (TSA, Difco, Becton  
90 Dickinson Co., Sparks, MD) plates.

91

92 **Susceptibility Testing.** MIC tests were performed in triplicate using broth microdilution in  
93 accordance with Clinical and Laboratory Standards Institute (CLSI) guidelines.(24, 25)  
94 The MIC was defined as the minimum concentration of antibiotic that will inhibit the  
95 visual growth of the isolated organism. Minimum bactericidal concentrations (MBC) were  
96 also determined in triplicate for each antimicrobial agent using CSLI guidelines.(25)  
97 Bacteria were quantified using CFU/mL, and 5 microliter aliquots were used for  
98 determination of MBC after 24 hours incubation at 37°C using TSA.(26)  
99  
100 **Endotracheal Tubes (ETTs).** Commercially available Sheridan® 6.0 mm ID, uncuffed  
101 ETTs (Hudson RIC, Temecula, CA, USA) were obtained. Each ETT was cut into 0.6 cm  
102 by 0.3 cm rectangular pieces (ETT chips) using a ¼ rectangle hand puncher (Fiskars  
103 Corporation, Helsinki, Finland), and sterilized with ethylene oxide gas prior to use in pre-  
104 formed and formed biofilm time kills experiments.(22) For comparison, we also tested  
105 commercially available PVC coupons (part Number RD 128-PVC, Biosurface  
106 Technologies, Corp, Bozeman, MT) for pre-formed biofilm *P. aeruginosa* PAO1.(16-18)  
107  
108 **Biofilm Formation.** Sterile ETT chips were placed in each well of a 24-well plate (BD  
109 Biosciences, San Jose, CA). The ETT chip was submerged with 2 mL of a final bacteria  
110 inoculum, either PAO1 or ATCC 25668, obtained as described above using TSB  
111 supplemented with 1% dextrose, 2% NaCl and 25 mg/L calcium (STSB) using modified  
112 Growing and Analyzing Static Biofilms.(27) The well plate was incubated at 37°C under  
113 static conditions for 24 hours to promote biofilm formation on ETT chips. After 24 hours,



114 each ETT chip was gently rinsed three times in sterile phosphate buffered saline (PBS)  
115 (Fisher Scientific, Pittsburgh, PA).

116

117 **Time Kill Study.** Using a 24h time kill study, three clinical conditions were modeled using  
118 *P. aeruginosa* strains PAO1 and 25668: 1) planktonic bacteria in the presence of the  
119 antibiotics tobramycin and polymyxin-E, without ETTs, 2) planktonic bacteria grown in  
120 the presence of *P. aeruginosa*, antibiotics and ETTs (simulating prevention) and 3) a 24h  
121 formed *P. aeruginosa* biofilm on ETTs prior to antibiotic exposure (simulating treatment).  
122 Each time kill experiment was carried out in a minimum of triplicate. All antimicrobial  
123 agents were tested at one, two, four and eight times their respective MIC with starting  
124 inocula of  $5.5\text{--}6.0 \times 10^6$  CFU/mL adjusted to McFarland standards using the Vitek  
125 colorimeter (bioMérieux, Inc, Durham, NC).(18, 28)

126

127 Sample aliquots (0.1 mL) were removed from cultures at 0, 4, and 24 hours after vortexing  
128 each tube for one minute to remove biofilm growth from the ETT chip.(22) Antimicrobial  
129 carryover was accounted for by serial dilution (10 – 10, 000 fold) of plated samples with  
130 normal saline or vacuum filtration. This methodology has a lower limit of detection of 2.0  
131  $\log_{10}$  CFU/mL.(28) Growth control tubes for each organism were prepared without  
132 antibiotic and run in parallel to the antibiotic test tubes.

133

134 For single antimicrobial agents, bactericidal activity (99.9% kill) was defined as a  $\geq 3$   
135  $\log_{10}$ CFU/mL reduction at 24h in colony count from the initial inoculum. Bacteriostatic  
136 activity was defined as a  $< 3 \log_{10}$  CFU/mL reduction at 24h in colony count from the

137 initial inoculum, while inactive was defined as no observed reduction from the initial  
138 inoculum.(23) For antibiotics evaluated in combination, synergy was defined as  $\geq 2 \log_{10}$   
139 CFU/mL decrease, indifference was defined as a 1 to 2  $\log_{10}$  CFU/mL change (increase or  
140 decrease), and antagonism was defined as  $>2 \log_{10}$  CFU/mL increase in growth compared  
141 to the most active single agent.

142

143 **Data analysis.** All statistical analyses were performed using SPSS statistical software  
144 (IBM SPSS statistics version 20, IBM Corporation, Armonk, NY USA). After 24 h of  
145 exposure to antimicrobial agent(s), the biofilm formation was quantified, bacteria at 4 hour  
146 and 24 hour were counted (with a lower limit of detection  $2.0 \log_{10}$  CFU/mL) to compare  
147 between antimicrobial groups, concentrations and strains using analysis of variance  
148 (ANOVA) followed by Tukey's *post-hoc* analysis. Multiple regressions for the association  
149 between substrates and CFU/mL were analyzed. A *p* value of  $\leq 0.05$  indicated statistical  
150 significance.

151 **RESULTS**

152 The MIC for tobramycin was 0.5 µg/mL and for polymyxin-E was 2 µg/mL for both PAO1 and  
153 25668 strains. The MBC for tobramycin was 4 and 32 µg/mL and for polymyxin-E was 16 and  
154 64 µg/mL respectively for Pseudomonas PAO1 and ATCC 25668 strains.

155

156 In the planktonic time kill study, tobramycin demonstrated bactericidal activity against both  
157 Pseudomonas isolates at 24 hours with average decrease of  $3.81 \pm 0.16 \log_{10}$  CFU/mL for all  
158 concentrations except 1 time the MIC for PAO1 (Fig. 1.a & b). Polymyxin-E demonstrated  
159 bacteriostatic activity at 2 and 4 times the MIC (average decrease of 2.16-2.63  $\log_{10}$ CFU/mL),  
160 and bactericidal activity at 8 times the MIC (average decrease of 3.07-3.56  $\log_{10}$ CFU/mL), but  
161 inactive at 1 time the MIC for both isolates at 24 hours (Fig. 1.c & d). The combination therapy  
162 at 2, 4, and 8 times the MIC demonstrated indifference with  $>3.44 \log_{10}$ CFU/mL kill for PAO1,  
163 and  $>3.46 \log_{10}$ CFU/mL kill for 25668 at 24 hours (Fig 1. e & f).

164

165 In the pre-formed biofilm time kill studies (simulating prevention) at 24 hours, tobramycin  
166 demonstrated bactericidal activity against both Pseudomonas isolates (average decrease of  $> 3.3$   
167  $\log_{10}$ CFU/mL), except 1 time the MIC for 25668 which showed inactivity ( $1.02 \pm 1.86$   
168  $\log_{10}$ CFU/mL increase; Fig 2.a & b). Similarly, polymyxin-E demonstrated bactericidal activity  
169 (average decrease of  $>3.08 \log_{10}$ CFU/mL) at greater than 2 times the MIC, but bacteriostatic  
170 activity at 1 time the MIC for both isolates at 24 hours (Fig. 2.c & d). Tobramycin and  
171 polymyxin-E combination demonstrated indifferent activity at all concentrations for both isolates  
172 (Fig. 2.e & f).

173

174 In formed biofilm time kill studies (simulating treatment) for PAO1, combination therapy at 4  
175 times the MIC was significantly more active at 4 hours compared to polymyxin-E alone at 4  
176 times (mean difference (MD)= -1.34, 95% confidence interval [CI], -2.4-0.3 log<sub>10</sub> CFU/mL, *p*=  
177 0.004) and 8 times (MD= -1.45, 95% CI, -2.5-0.4 log<sub>10</sub> CFU/mL, *p*=0.001) the MIC. Similarly,  
178 combination therapy at 8 times the MIC was significantly more active at 4 hours compared to  
179 polymyxin-E alone at 8 times the MIC (MD= -1.23, 95% CI, -2.3-0.2 log<sub>10</sub> CFU/mL, *p*=0.001).  
180 However, indifferent activity was observed at 24 hours. Similarly, for 25668, combination  
181 therapy at 8 times the MIC was significantly more active compared to polymyxin-E alone at 4  
182 hours (MD= -1.06, 95% CI, -1.7-0.4 log<sub>10</sub> CFU/mL, *p*< 0.001). However, indifferent activity  
183 was observed at 24 hours. Once biofilm is formed, both single agent and combination antibiotics  
184 resulted in inactivity or indifference (Fig. 3. a - f).

185

186 In addition to medical grade PVC ETTs, we assayed time kill using commercially available PVC  
187 coupons(16-18). A similar trend of bactericidal activity was demonstrated at 24 hours with  
188 greater than 4 times the MIC of tobramycin (average decrease of >3.03 log<sub>10</sub> CFU/mL) and with  
189 greater than 2 times the MIC of polymyxin-E (average decrease of >3.1 log<sub>10</sub> CFU/mL), but  
190 indifference was noted when the combination of tobramycin and polymyxin-E was evaluated at  
191 2, 4, and 8 times the MIC (average decrease of > 3.21 log<sub>10</sub>CFU/mL). ANOVA showed that  
192 there was a significant difference between substrates and CFU/mL at 4 hours (MD= 0.08, 95%  
193 CI, 2.5-3.7 log<sub>10</sub> CFU/mL, *p*= 0.041) (Table 1). Multiple regression analysis demonstrated that  
194 there was a significant association between CFU/mL with substrate at 4 hours (partial eta  
195 squared [eta] =0.493, *p*< 0.001) and at 24 hours (eta=0.208, *p*<0.001). The overall model fit was  
196  $R^2 = 0.954$ .

197

198 **DISCUSSION**

199 Ventilator associated pneumonia, a common nosocomial infection often caused by bacteria that  
200 produce biofilm, results in increased morbidity, medical costs and multi-drug resistant  
201 organisms.(2, 3, 29-32) In one study, adult patients with VAP were hospitalized longer (38 vs.  
202 13 days,  $p<0.01$ ), mortality rates were higher (50% vs. 34%,  $p<0.01$ ), and hospital costs were  
203 greater (\$70,568 vs. \$21,620,  $p< 0.01$ ) compared to uninfected ventilated patients, with estimated  
204 VAP attributable costs of \$11,897.(32) However, limited diagnostic criteria and modification of  
205 ETTs make VAP prevention particularly challenging and difficult especially for neonates and  
206 children.(2)

207 In children, re-intubation and tracheostomy insertion create the additional risk of damaging their  
208 small and fragile airway; therefore, re-intubation or tracheostomy after a standard duration of  
209 intubation is not routinely practiced. Thus, the longer the ETTs remain in patients due to  
210 prolonged mechanical ventilation, the more likely biofilms are to develop and adhere.(33-35)  
211 This bacterial accumulation of biofilms on ETTs may become dislodged during simple routine  
212 care such as suctioning or due to ventilation air flow. Bacteria and biofilm that break off become  
213 planktonic and seed further in the airway, causing more complicated pneumonia.(8, 36)

214  
215 One controversial approach to treatment of VAP is “selective decontamination of the digestive  
216 tract” with broad spectrum intravenous (IV) antimicrobials.(37, 38) However, IV prophylaxis is  
217 not widely accepted due to fear of creating antibiotic resistant strains among VAP pathogens. In  
218 the pediatric population, one of the most common VAP pathogen is *P. aeruginosa*, accounting  
219 for 17-25% of VAP.(2, 4, 5) Our model is most consistent with the practice of direct instillation  
220 of liquid antimicrobial agents through the ETT as prophylaxis against, or treatment of VAP  
221 caused by *P. aeruginosa* compared to inhalation of nebulized antibiotics.(39) Instillation

222 treatments pose less risk of systemic toxicity than IV administration because antimicrobial agents  
223 can be delivered locally using ETTs or tracheostomy tubes in children and neonates. Moreover,  
224 instillation can deliver drug directly to the site of pneumonia whereas nebulized drug may adsorb  
225 on the ETT, permeate into the ETT wall, or remain in the proximal airway. Therefore, our study  
226 model using ETT chips is useful to help understand the effects of tobramycin and polymyxin-E,  
227 alone or in combination, to treat VAP cause by *P. aeruginosa*.

228

229 In our study, we examined *P. aeruginosa* growth with or without the presence of medical grade  
230 polyvinyl chloride (PVC) ETT to evaluate the bactericidal effects of two antibiotics in the  
231 condition of VAP. We found that in an in vitro condition, the bactericidal effect of tobramycin  
232 or polymyxin-E monotherapy required greater than 2 times the MIC at 24 hours for the pre-  
233 biofilm condition (prevention). However, antibiotics demonstrated different activity against the  
234 two different strains. For PAO1, tobramycin monotherapy was equally active for killing  
235 compared to the combination approach. For 25668, the combination therapy was more active  
236 for killing compared to monotherapy at 24 hours (Fig 2.a-f); this finding may be related to the  
237 biofilm-forming abilities of each bacterium.

238

239 Our study also demonstrated that two of the antibiotics tested in either monotherapy or in  
240 combination showed inactivity or indifference once biofilm was formed on ETTs against both  
241 *Pseudomonas* strains (Fig 3.a - f). This is in contrast to the conclusions drawn by Herrmann et  
242 al. using a 96-peg Calgary biofilm device in vitro showing combination therapy with colistin-  
243 tobramycin combination was superior to monotherapy against *Pseudomonas* biofilm.(40)

244

245 Many in vitro studies have used commercially available PVC coupons, which have different  
246 texture and flexibility (based on the plasticizer content compared to medical grade PVC ETTs).  
247 We hypothesized that bacterial colonies would form differently on commercially available PVC  
248 coupons compared to medical grade PVC ETTs. To capture *Pseudomonas* growth in relation to  
249 different material surfaces more accurately, we studied the same antibiotic therapy against  
250 *Pseudomonas* PAO1 using both PVC coupon and PVC ETTs. There was the significant  
251 association among CFU/mL and substrate at 4 and 24 hours (Table 1), thus it showed the  
252 importance of utilizing same device material to mimic VAP condition to evaluate antibiotic  
253 activity on biofilm.

254

255 In conclusion, neither single nor combination therapy with tobramycin and/or polymyxin-E  
256 demonstrated killing activity once *Pseudomonas* biofilm was already formed on ETTs, however,  
257 no antagonism was noted. Bactericidal effects against pre-formed biofilm (simulating  
258 prevention) in the presence of ETTs suggest that surveillance cultures could identify pathogens  
259 prior to biofilm formation, and allow prophylactic or targeted approaches to therapy, especially  
260 when *Pseudomonas* is a potential pathogen. In addition, this study demonstrated the importance  
261 of material choice in vitro time kill study. Further investigation could incorporate wild type  
262 strains as well as clinically feasible treatment options for VAP in children.

263 **ACKNOWLEDGEMENTS**

264 This study was funded by PCCSDP (Pediatric Critical Care Scientist Development Program),  
265 through NIH 5K12HD047349-07, and supported by Department of Pediatrics, Hasbro Children's  
266 Hospital, Rhode Island Hospital.

267 Presented as a poster (E-790) at the Interscience Conference of Antimicrobial Agents and  
268 Chemotherapy 2012 meeting, San Francisco CA, September 7-12th 2012.

269



270 **REFERENCES**

- 271 1. **Srinivasan R, Asselin J, Gildengorin G, Wiener-Kronish J, Flori HR.** 2009. A  
272 prospective study of ventilator-associated pneumonia in children. *Pediatrics* **123**:1108-  
273 1115.
- 274 2. **Foglia E, Meier MD, Elward A.** 2007. Ventilator-associated pneumonia in neonatal and  
275 pediatric intensive care unit patients. *Clin Microbiol Rev* **20**:409-425, table of contents.
- 276 3. **Stockwell J A.** 2007. Nosocomial infections in the pediatric intensive care unit: affecting  
277 the impact on safety and outcome. *Pediatr Crit Care Med* **8**:S21-37.
- 278 4. **Weber D J, Rutala WA, Sickbert-Bennett EE, Samsa GP, Brown V, Niederman MS.**  
279 2007. Microbiology of ventilator-associated pneumonia compared with that of hospital-  
280 acquired pneumonia. *Infect Control Hosp Epidemiol* **28**:825-831.
- 281 5. **Babcock H, Zack JE, Garrison T, Trovillion E, Kollef MH, Fraser VJ.** 2003.  
282 Ventilator-associated pneumonia in a multi-hospital system: differences in microbiology  
283 by location. *Infect Control Hosp Epidemiol* **24**:853-858.
- 284 6. **Costerton J, Montanaro L, Arciola CR.** 2005. Biofilm in implant infections: its  
285 production and regulation. *Int J Artif Organs* **28**:1062-1068.
- 286 7. **Talsma S S.** 2007. Biofilms on medical devices. *Home Healthc Nurse* **25**:589-594.
- 287 8. **Adair C, Gorman SP, Feron BM, Byers LM, Jones DS, Goldsmith CE, Moore JE,**  
288 **Kerr JR, Curran MD, Hogg G, Webb CH, McCarthy GJ, Milligan KR.** 1999.  
289 Implications of endotracheal tube biofilm for ventilator-associated pneumonia. *Intensive*  
290 *Care Med* **25**:1072-1076.
- 291 9. **Estes R JG U Meduri.** 1995. The pathogenesis of ventilator-associated pneumonia: I.  
292 Mechanisms of bacterial transcolonization and airway inoculation. *Intensive Care Med*  
293 **21**:365-383.







- 294 10. **Gil-Perotin S, P Ramirez, V Marti, J M Sahuquillo, E Gonzalez, I Calleja, R**  
295 **MenendezJ Bonastre.** 2012. Implications of endotracheal tube biofilm in ventilator-  
296 associated pneumonia response: a state of concept. Crit Care **16**:R93.
- 297 11. **Inglis T J, M R Millar, J G JonesD A Robinson.** 1989. Tracheal tube biofilm as a  
298 source of bacterial colonization of the lung. J Clin Microbiol **27**:2014-2018.
- 299 12. **Gorman S, McGovern JG, Woolfson AD, Adair CG, Jones DS.** 2001. The  
300 concomitant development of poly(vinyl chloride)-related biofilm and antimicrobial  
301 resistance in relation to ventilator-associated pneumonia. Biomaterials **22**:2741-2747.
- 302 13. **Zavascki A, Li J, Nation RL, Superti SV, Barth AL, Lutz L, Ramos F, Boniatti MM,**  
303 **Goldani LZ.** 2009. Stable polymyxin B susceptibility to *Pseudomonas aeruginosa* and  
304 *Acinetobacter* spp. despite persistent recovery of these organisms from respiratory  
305 secretions of patients with ventilator-associated pneumonia treated with this drug. J Clin  
306 Microbiol **47**:3064-3065.
- 307 14. **Goyal K, Gautam V, Ray P.** 2012. Doripenem vs meropenem against *Pseudomonas* and  
308 *Acinetobacter*. Indian J Med Microbiol **30**:350-351.
- 309 15. **Kiem S, Schentag JJ.** 2006. Relationship of minimal inhibitory concentration and  
310 bactericidal activity to efficacy of antibiotics for treatment of ventilator-associated  
311 pneumonia. Semin Respir Crit Care Med **27**:51-67.
- 312 16. **Ceri H, Olson ME, Stremick C, Read RR, Morck D, Buret A.** 1999. The Calgary  
313 Biofilm Device: new technology for rapid determination of antibiotic susceptibilities of  
314 bacterial biofilms. J Clin Microbiol **37**:1771-1776.
- 315 17. **Garey K, Vo QP, Lewis RE, Saengcharoen W, LaRocco MT, Tam VH.** 2009.  
316 Increased bacterial adherence and biomass in *Pseudomonas aeruginosa* bacteria exposed  
317 to clarithromycin. Diagn Microbiol Infect Dis **63**:81-86.

- 318 18. **Hadi R, Vickery K, Deva A, Charlton T.** 2010. Biofilm removal by medical device  
319 cleaners: comparison of two bioreactor detection assays. *J Hosp Infect* **74**:160-167.
- 320 19. **Murray T, Kazmierczak BI.** 2008. *Pseudomonas aeruginosa* exhibits sliding motility in  
321 the absence of type IV pili and flagella. *J Bacteriol* **190**:2700-2708.
- 322 20. **Murray T, Okegbe C, Gao Y, Kazmierczak BI, Motterlini R, Dietrich LE, Bruscia**  
323 **EM.** 2012. The carbon monoxide releasing molecule CORM-2 attenuates *Pseudomonas*  
324 *aeruginosa* biofilm formation. *PLoS One* **7**:e35499.
- 325 21. **Klausen M, Heydorn A, Ragas P, Lambertsen L, Aaes-Jørgensen A, Molin S,**  
326 **Tolker-Nielsen T.** 2003. Biofilm formation by *Pseudomonas aeruginosa* wild type,  
327 flagella and type IV pili mutants. *Molecular Microbiology* **48**:1511-1524.
- 328 22. **Seil J T, Rubien NM, Webster TJ, Tarquinio KM.** 2011. Comparison of quantification  
329 methods illustrates reduced *Pseudomonas aeruginosa* activity on nanorough polyvinyl  
330 chloride. *Journal of Biomedical Materials Research Part B: Applied Biomaterials* **98B**:1-  
331 7.
- 332 23. **LaPlante K, Sakoulas G.** 2009. Evaluating aztreonam and ceftazidime  
333 pharmacodynamics with *Escherichia coli* in combination with daptomycin, linezolid, or  
334 vancomycin in an in vitro pharmacodynamic model. *Antimicrob Agents Chemother*  
335 **53**:4549-4555.
- 336 24. **Clinical and Laboratory Standards Institute.** 2013. Performance Standards for  
337 Antimicrobial Susceptibility Testing M100-S23, 23rd Informational Supplement ed.  
338 Clinical and Laboratory Standards Institute, Wayne PA.
- 339 25. **Clinical and Laboratory Standards Institute.** 1999. Methods for Determining  
340 Bactericidal Activity of Antimicrobial Agents; Approval Guideline M26-A, First ed.  
341 Clinical and Laboratory Standards Institute, Wayne, PA.

- 342 26. **LaPlante K, Rybak MJ.** 2004. Impact of high-inoculum *Staphylococcus aureus* on the  
343 activities of nafcillin, vancomycin, linezolid, and daptomycin, alone and in combination  
344 with gentamicin, in an in vitro pharmacodynamic model. *Antimicrob Agents Chemother*  
345 **48**:4665-4672.
- 346 27. **Merritt J, Kadouri DE, O'Toole GA** 2005. Growing and analyzing static biofilms.  
347 *Current Protocols in Microbiology* **Chapter 1: Unit 1B.1.:1.1.**
- 348 28. **LaPlante K, Rybak MJ.** 2004. Clinical glycopeptide-intermediate staphylococci tested  
349 against arbekacin, daptomycin, and tigecycline. *Diagn Microbiol Infect Dis* **50**:125-130.
- 350 29. **Chastre J, Fagon JY.** 2002. Ventilator-associated pneumonia. *Am J Respir Crit Care*  
351 *Med* **165**:867-903.
- 352 30. **Fagon J, Chastre J, Domart Y, Trouillet JL, Gibert C.** 1996. Mortality due to  
353 ventilator-associated pneumonia or colonization with *Pseudomonas* or *Acinetobacter*  
354 species: assessment by quantitative culture of samples obtained by a protected specimen  
355 brush. *Clin Infect Dis* **23**:538-542.
- 356 31. **Elward A, Warren DK, Fraser VJ.** 2002. Ventilator-associated pneumonia in pediatric  
357 intensive care unit patients: risk factors and outcomes. *Pediatrics* **109**:758-764.
- 358 32. **Warren D, Shukla SJ, Olsen MA, Kollef MH, Hollenbeak CS, Cox MJ, Cohen MM,**  
359 **Fraser VJ.** 2003. Outcome and attributable cost of ventilator-associated pneumonia  
360 among intensive care unit patients in a suburban medical center. *Crit Care Med* **31**:1312-  
361 1317.
- 362 33. **Levine S AM S Niederman.** 1991. The impact of tracheal intubation on host defenses  
363 and risks for nosocomial pneumonia. *Clin Chest Med* **12**:523-543.

- 364 34. **Feldman C, M Kassel, J Cantrell, S Kaka, R Morar, A Goolam Mahomed, J I Philips.**  
365 1999. The presence and sequence of endotracheal tube colonization in patients  
366 undergoing mechanical ventilation. *Eur Respir J* **13**:546-551.
- 367 35. **Gibbs KI R Holzman.** 2012. Endotracheal tube: friend or foe? Bacteria, the endotracheal  
368 tube, and the impact of colonization and infection. *Semin Perinatol* **36**:454-461.
- 369 36. **Bauer T, Torres A, Ferrer R, Heyer CM, Schultze-Werninghaus G, Rasche K.** 2002.  
370 Biofilm formation in endotracheal tubes. Association between pneumonia and the  
371 persistence of pathogens. *Monaldi Arch Chest Dis* **57**:84-87.
- 372 37. **Bonten M J.** 2002. Strategies for prevention of hospital-acquired pneumonia: oral and  
373 selective decontamination of the gastrointestinal tract. *Semin Respir Crit Care Med*  
374 **23**:481-488.
- 375 38. **van Essen E, de Jonge E.** 2011. Selective decontamination of the digestive tract (SDD):  
376 is the game worth the candle? *Semin Respir Crit Care Med* **32**:236-242.
- 377 39. **Brown R B, J A Kruse, G W Counts, J A Russell, N V Christou, M L Sands.** 1990.  
378 Double-blind study of endotracheal tobramycin in the treatment of gram-negative  
379 bacterial pneumonia. The Endotracheal Tobramycin Study Group. *Antimicrob Agents*  
380 *Chemother* **34**:269-272.
- 381 40. **Herrmann G, Yang L, Wu H, Song Z, Wang H, Hoiby N, Ulrich M, Molin S,**  
382 **Riethmuller J, Doring G.** 2010. Colistin-tobramycin combinations are superior to  
383 monotherapy concerning the killing of biofilm *Pseudomonas aeruginosa*. *J Infect Dis*  
384 **202**:1585-1592.
- 385  
386

387 FIGURES



389 **Figure 1.a-f.** Time Kill against planktonic *P. aeruginosa*. Tobramycin against planktonic *P.*  
390 *aeruginosa* PAO1 (Fig 1.a), 25668 (Fig 1.b), polymyxin-E against planktonic *P. aeruginosa*  
391 PAO1 (Fig 1.c), 25668 (Fig 1.d), combination of tobramycin and polymyxin-E against  
392 planktonic *P. aeruginosa* PAO1 (Fig 1.e) and 25668 (Fig 1.f). Results are presented as mean  $\pm$   
393 standard deviation.

Fig. 2. a




Fig. 2. b




Fig. 2. c




Fig. 2. d




Fig. 2. e





Fig. 2. f





395 **Figure 2.a-f.** Time Kill against pre-biofilm formed *P. aeruginosa*. Tobramycin against *P.*  
396 *aeruginosa* PAO1 (Fig 2.a), 25668 (Fig 2.b), polymyxin-E against *P. aeruginosa* PAO1 (Fig  
397 2.c), 25668 (Fig 2.d), combination of tobramycin and polymyxin-E against *P. aeruginosa* PAO1  
398 (Fig 2.e) and 25668 (Fig 2.f) with presence of ETT chips. Results are presented as mean  $\pm$   
399 standard deviation.

Fig. 3. a




Fig. 3. b




Fig. 3. c




Fig. 3. d




Fig. 3. e





Fig. 3. f



401 **Figure 3. a-f.** Time Kill against biofilm formed *P. aeruginosa*. Tobramycin against *P.*  
402 *aeruginosa* PAO1 (Fig 3.a), 25668 (Fig 3.b), polymyxin-E against *P. aeruginosa* PAO1 (Fig  
403 3.c), 25668 (Fig 3.d), combination of tobramycin and polymyxin-E against *P. aeruginosa* PAO1  
404 (Fig 3.e) and 25668 (Fig 3.f). Results are presented as mean  $\pm$  standard deviation.

405 **TABLE**

406 **Table 1:** Comparison of log<sub>10</sub> colony forming unit/mL (CFU/mL) change at 4 hours from 0 hr  
 407 growth control between endotracheal tube (ETT) and polyvinyl chloride (PVC) coupons.

	Tobramycin		Polymyxin-E		Tobramycin + Polymyxin-E	
	ETT	PVC coupon	ETT	PVC coupon	ETT	PVC coupon
1 x MIC	-2.81± 0.04	-2.82±0.12	-3.44±0.43	-1.77±0.04	-3.36± 0.46	-4.09±0.10
2 x MIC	-3.97 ±0.21	-0.71±0.01	-3.92± 0.03	-2.86±0.10	-3.99±0.10	-4.24±0.03
4 x MIC	-3.97 ±0.21	-1.23±0.03	-3.92±0.03	-3.16±1.2	-3.99±0.10	-4.24±0.03
8 x MIC	-3.97 ±0.21	-1.06±0.03	-3.92±0.03	-3.49±1.28	-3.99±0.10	-4.24±0.03

408 Average change ± standard deviation.