2015

PhonoSys: Mobile Phonocardiography Diagnostic System for Newborns

Amir Mohammad Amiri
University of Rhode Island

Giuliano Armano

Amir Mohammad Rahmani

Kunal Mankodiya
University of Rhode Island, kunalm@uri.edu

Follow this and additional works at: https://digitalcommons.uri.edu/ele_facpubs

The University of Rhode Island Faculty have made this article openly available.
Please let us know how Open Access to this research benefits you.

Terms of Use

This article is made available under the terms and conditions applicable towards Open Access Policy Articles, as set forth in our [Terms of Use](#).

Citation/Publisher Attribution

Amiri, A. M., Armano, G., Rahmani, A. M., & Mankodiya, K. (2015, October 14-16). *PhonoSys: Mobile Phonocardiography Diagnostic System for Newborns*. Proceedings of the 5th EAI International Conference on Wireless Mobile Communication and Healthcare, London, Great Britain. doi: 10.4108/eai.14-10-2015.2261614

Available at: <http://dx.doi.org/10.4108/eai.14-10-2015.2261614>

This Conference Proceeding is brought to you for free and open access by the Department of Electrical, Computer, and Biomedical Engineering at DigitalCommons@URI. It has been accepted for inclusion in Department of Electrical, Computer, and Biomedical Engineering Faculty Publications by an authorized administrator of DigitalCommons@URI. For more information, please contact digitalcommons-group@uri.edu.

PhonoSys: Mobile Phonocardiography Diagnostic System for Newborns *

Amir Mohammad Amiri,[†]
Kunal Mankodiya
Department of Electrical,
Computer, and Biomedical
Engineering
University of Rhode Island
RI, USA
amiri,kunalm@ele.uri.edu

Giuliano Armano
Department of Electrical and
Electronic Engineering (DIEE)
University of Cagliari
Cagliari, Italy
armano@diee.unica.it

Amir Mohammad
Rahmani
Department of Information
Technology
University of Turku
Turku, Finland
amirah@utu.fi

ABSTRACT

Heart murmurs have been found to be a life threatening condition for the newborns who are born with cardiac abnormalities. The first sign of pathological changes of heart valves appear in phonocardiogram (PCG) which contains very useful information about cardiovascular system. It is a challenging venture to distinguish pathological murmurs from innocent ones. In this paper, we have developed a diagnostic algorithm called PhonoSys to analyze PCG using random forest. PhonoSys algorithm will run on mobile devices for remote PCG analysis. We recorded PCG signals from 120 newborns who are either healthy or with cardiac abnormalities. Eventually, in this study, 97.6% accuracy, 96.8% sensitivity, and 98.4% specificity were obtained to classify between innocent and pathological murmurs.

Keywords

Phonocardiogram, m-Health, Random forest, Newborn.

1. INTRODUCTION

Heart murmur have been reported in 0.3-77.4% of newborns, with prevalence dependent on several factors, innocent or pathological. S. B. Ainsworth et al published study of heart murmurs in 1999, which indicates 46% has innocent murmurs and 54% of newborns with murmurs are underling pathological case [1].

PCG contains very important information of cardiovascular system which is widely used by physicians to evaluate cardiac functions in patients and detect the presence of abnormalities. PCG is produced by vibrations of heart and

blood circulation which creates two major sounds (S1 and S2) and heart murmurs. A murmur that caused by contracting of the heart muscle is called systolic murmur which is located between S1 and S2. A murmur that occurs when the heart muscle relaxes between beats is called a diastolic murmur which is located between S2 and S1. Nowadays signals produced by the heart are not only heard using a stethoscope but also observed as PCG on a screen (see figure 1).

Figure 1: Sample of two-cycle PCG, where components S1, S2 and heart murmurs are highlighted.

Newborns often have heart murmurs that mostly originate from normal flow patterns with no structural or anatomic abnormalities of the heart or vessels. The type-I of murmur referred to as innocent, physiological or normal murmurs. For instance the most common cause of murmur in newborns is when a specific condition called patent ductus arteriosus (PDA) occurs; it is often detected shortly after birth, most commonly in premature newborns [7]. PDA is a potentially serious condition in which blood circulates abnormally between two of the major arteries near the heart, due to the failure of a blood vessel between these arteries to properly close. In most cases, the only symptom of PDA is a heart murmur until the ductus closes on its own shortly after birth [3]. If the murmur is still present at 2 weeks, there is a possibility of murmur type-II and paediatrician may refer newborn to the cardiologist.

The second type of murmurs are caused by a problem in structure of the heart and is called pathological or abnormal heart murmurs. A problem with the structure of the

*(Produces the permission block, and copyright information). For use with SIG-ALTERNATE.CLS. Supported by ACM.

[†]Dr. Amiri insisted his name be first.

Permission to make digital or hard copies of all or part of this work for personal or classroom use is granted without fee provided that copies are not made or distributed for profit or commercial advantage and that copies bear this notice and the full citation on the first page. To copy otherwise, to republish, to post on servers or to redistribute to lists, requires prior specific permission and/or a fee.

WOODSTOCK '97 El Paso, Texas USA

Copyright 20XX ACM X-XXXXX-XX-X/XX/XX ...\$15.00.

heart that is present at birth is called Congenital Heart Defect (CHD) [12]. CHDs are the most common type of birth defect.

Some babies born with a CHD can appear healthy at first and can be sent home with their family before their heart defect is treated or needs a regular health care. CHDs affect approximately 1 in 125 live births [10]. Of these, 30% have extra-cardiac anomalies (such as tracheoesophageal fistula, anorectal anomalies), which might require surgery within the first year of life [13]. These babies are at risk for having serious problems within the first few days or weeks of life and often require emergency care.

It is a challenging venture to distinguish pathological murmurs from innocent ones. In this paper we have developed a diagnostic algorithm called PhonoSys to analyze PCG using random forest.

2. RELATED WORK

The basic technique by S. Yuenyong [14], which reports an automatic tele-auscultation system. Segmentation of the heart sound is performed using ECG signal as reference. Feature vector is constructed for each cardiac cycle by analyzing the systole and diastole of each cycle by applying discrete wavelet transform. Features were classified using pattern recognition neural network.

W. Jin-gang [9] introduces a remote monitoring system for PCG is constructed, which integrates embedded Internet technology and wireless technology.

T. Chen [6] reports preliminary work performed on a gold standard database and a cellphone platform. Results indicate that HR and HRV can be accurately assessed from PCG using only a cellphone and hands-free kit. Heart sound analysis software, which can run on a standard cellphone in real time, has been developed that detects S1 heart sounds with a sensitivity of 92.1% and a positive predictivity of 88.4%. But there are a very few works that allow extracting pathological information from PCG. In this paper, we demonstrate an approach to classify clinically-relevant, pathological data from PCG on mobile devices.

3. METHODS

The PhonoSys algorithm consists of four major steps: (A) Data acquisition; (B) Pre-processing; (C) Feature extraction and (D) Classification (see figure 2).

Figure 2: PhonoSys implementation scheme.

3.1 Data acquisition

PCG data has recorded by an electronic stethoscope which was connected to a mobile. We recorded the PCG data from

120 newborns who were 1-20 old at Imam Ghaem Hospital, Mashhad, Iran. All newborns were visited by a cardiologist who used echocardiography to label the PCG data with information on murmur no murmur. The label for specifying it that murmur is innocent or pathological. We received consent of all parents.

3.2 Pre-processing

Data has recorded in many situations in hospital and clinic that the signal is corrupted by different types of noise, such as power line, electromyographic, respiration and microphone movement artifact, which can affect on PCG signals and the algorithm. The preprocessing stage improves the classification accuracy of any algorithm; because, it gives us more accurate features.

Figure 3 shows a spectrogram of PCG signal. Since the main spectrum of PCG occurs within the range of 150 Hz, the system filters the original PCG using a 3rd order band-pass Butterworth *filter*, with cut-off frequencies at 50-200 Hz.

Figure 3: PCG spectrogram.

The second step of preprocessing is *segmentation*. We devised and implemented an algorithm aimed at identifying S1 and S2 based on the timing between high-amplitude components. To this end, we used the *Gabor Wavelet* for peak detection (see [8]), which can be formally described as follows:

$$\Psi(t) = C \cdot e^{-j\omega t} \cdot e^{-t^2} \quad (1)$$

where $e^{-j\omega t} \cdot e^{-t^2}$ is the complex Gaussian function and C is a normalizing constant. The threshold was used to identify the peaks. The *threshold* value was set to 0.1 for wavelet scale coefficients (PCG components).

Latter step is *zero-crossing* which has used to find the spots where peaks occur, the number of zero crossings per segment being also an equivalent representation of the dominant component of a signal segment. The algorithm calculates the size of the intervals in which the value of the function is zero (systolic and diastolic). Let us recall that systolic (S1-S2) and diastolic (S2-S1) murmurs occur respectively in the smaller and bigger time interval (see Figure 4).

After segmentation for detecting PCG components and heart murmurs (systolic and diastolic murmurs), it is very important to select a cycle instead of whole signal which can reduce the number of data and the cost of computing. The proposed algorithm to find best cycle is based on which cycle shows more signs of heart disease –as detecting the most informative cycle can optimize the performance of the classification process. For each patient, recorded data include

Figure 4: PCG segmentation to detect systolic and diastolic area; a) Raw signal, b) Complex Gaussian Wavelet coefficients.

several cardiac cycles on a time span of few seconds. Despite the fact that filtering has been implemented to remove noise, the residual noise may be part of the heart sound signal –such as respiratory sound, artifact noise or newborn voices. Pearson’s Correlation Coefficient (PCC) has been used to select the cycle with minimum noise and most properties of the whole signal. PCC between two signals X and Y (cardiac cycles, in this context) is defined as follows:

$$r_{XY} = \frac{SS_{XY}}{\sqrt{(SS_{XX})(SS_{YY})}} \quad (2)$$

where:¹

- $SS_{XY} = \sum XY - \frac{1}{n} \cdot \sum X \cdot \sum Y$
- $SS_{XX} = \sum X^2 - \frac{1}{n} \cdot (\sum X)^2$
- $SS_{YY} = \sum Y^2 - \frac{1}{n} \cdot (\sum Y)^2$

The Pearson correlation coefficient is calculated between pairs of signals, each signal including several cycles. An example of correlation is shown in Table 3.2. The overall correlation for each cardiac cycle $C_{i,j}$ is obtained through the following formula:

$$r_{C_i} = \frac{1}{n-1} \sum_{i \neq j}^j C_{i,j} \quad (3)$$

where n is the number of cardiac cycles. In the given example C_4 and C_2 are selected as best and worst cycles according to Equation (3) which means the cycle with minimum and maximum noise (and hence with the most informative content for the whole signal).

3.3 Feature extraction

This phase is focused on extracting features of the signal that better highlight characteristic properties of the PCG signal, with the goal of identifying those that are more suitable for the classification purpose. The features extraction are proposed for two goals, heart murmurs detection and diagnosis.

Different signal processing tools are used on smart-phone or tablets for detecting heart murmurs. For instance, maximum and minimum value amplitude, total Absolute Area,

¹ n is number of data pairs for each sample; in this case 1500.

Table 1: The average coefficients of correlation between cycles.

	C_1	C_2	C_3	C_4
C_1	1	-0.15	0.13	0.16
C_2	-0.15	1	-0.99	-0.51
C_3	0.13	-0.99	1	0.85
C_4	0.16	-0.51	0.85	1
$\Sigma C_{i,j}$	1.14	-0.65	0.99	1.5

peak to peak time window and variance are some of these signal processing tools that are used on smart phone or tablet. Based on our research a range value has defined for each categorizes. The smart-phone display that the newborn has heart murmurs or no. If yes, smart-phone send selected cycle to cloud for future computing. Is that an innocent or pathological murmur? The algorithm on smart-phone is defined as:

- Recording or receiving PCG signals.
- Filtering for removing unwanted noises.
- Segment signal into systolic and diastolic murmurs.
- Data reduction by selecting the best cardiac cycle.
- Features extraction to detect heart murmurs, Is there any heart murmurs?
 1. Yes, Send selected cardiac cycle for heart murmurs diagnosis (innocent or pathological murmur)
 2. No, display data as normal and without heart murmurs

The latter step was aimed at, extracting features according to distinguish between innocent and pathological murmurs. Features are extracted on the basis of signal information in alternative domains, such as time, frequency, or time-frequency domains. The features with respect to the templates were Shannon Energy, Bispectrum, Wigner Distribution and Wigner Bispectrum [2].

3.4 Classification

A novel application of random forest for clinical and physiological data in monitoring and remote diagnosis is presented in this work.

The random forest methodology, developed by Leo Breiman [5] in 2001. Random forest has been proved an excellent performance in classification tasks which is a general class of ensemble building methods using a decision tree as the base classifier [11]. Random forest is an ensemble learning algorithm that composed of a collection of individual tree-structured classifiers, and the output is based on majority vote among the random tree classifiers (see figure 5).

The learning algorithm choose their splitting features from a random subset of k features at each internal node. Randomness is injected into each tree in two ways, the utilize of bootstrapping to sample from the original dataset or by growing each tree on different random subsamples and determining splitter partly at random [6, 4].

In the each decision tree for predicting heart murmurs, we start at the top of the tree and follow different branches, depending on conditions involving the predictor variables.

Figure 5: Structure of the random forest classifier.

Once we arrive at an end-point of the tree, we used 50 trees and variables classified in two classes (0=innocent and 1=pathological murmurs).

4. EXPERIMENTS AND RESULTS

The likelihood ratio (LR) is calculated to archiving sensitivity and specificity on a random forest, define into positive and negative ratio as follows:

$$LR+ = \frac{\text{sensitivity}}{1 - \text{specificity}}, LR- = \frac{1 - \text{sensitivity}}{\text{specificity}} \quad (4)$$

Likelihood ratio positive is the probability for a newborn that has pathological murmur of being classified as unhealthy, divided by the probability for a newborn that given test result would be expected as healthy.

Likelihood ratio negative is the probability for a newborn that has innocent murmur of being classified as healthy, divided by the probability for a newborn that given test result would be expected as unhealthy.

Table 2 in the form of a confusion matrix presents the results for real-time classification of PCG data with accuracy percentage. Experiments have been performed on a balanced set of 120 samples (meaning that the number of samples was the same for innocent and pathological murmurs).

Table 2: Classification result of heart murmurs in newborns.

	Innocent	Pathological	Correct%
Innocent	59 (98.4%)	1 (1.6%)	98.4%
Pathological	2 (3.2%)	58 (96.8%)	96.8%
Average/Overall	120		97.6%

It can be seen that out of 60 PCG signal with innocent murmur, 59 were recognized as healthy and 1.6% were not recognized. Similarly, out of 60 PCG signal with pathological murmur, 58 were recognized as pathological and 2 were not recognized. Overall sensitivity for pathological murmur was 96.8% with a specificity of 98.4%.

5. CONCLUSION

PhonoSys is an mHealth system for monitoring and remote diagnosis of heart murmurs using PCG signal. We provided an implementation of algorithm framework for monitoring PCG signal and a remote diagnosis of heart murmurs

on mobile devices. This methods includes novelties in both, reducing amount of data by selecting best cycle, and applying random forest algorithm. The diagnostic system shows a high fidelity to distinguishing between innocent and pathological murmurs. The proposed system will aid physicians in remote monitoring and diagnosis of patients using mobile device and/or wearable sensors.

6. REFERENCES

- [1] S. B. Ainsworth, J. P. Wyllie, and C. Wren. Prevalence and clinical significance of cardiac murmurs in neonates. *Archives of Disease in Childhood-Fetal and Neonatal Edition*, 80(1):F43–F45, 1999.
- [2] A. M. Amiri. *An intelligent diagnostic system for screening newborns*. PhD thesis, Universita’ degli Studi di Cagliari, Italy, 5 2015.
- [3] A. M. Amiri and G. Armano. Segmentation and feature extraction of heart murmurs in newborns. *Journal of Life Sciences and Technologies Vol*, 1(2), 2013.
- [4] G. Biau, L. Devroye, and G. Lugosi. Consistency of random forests and other averaging classifiers. *The Journal of Machine Learning Research*, 9:2015–2033, 2008.
- [5] L. Breiman. Random forests. *Machine learning*, 45(1):5–32, 2001.
- [6] T. Chen, K. Kuan, L. A. Celi, and G. D. Clifford. Intelligent heartsound diagnostics on a cellphone using a hands-free kit. In *AAAI Spring Symposium: Artificial Intelligence for Development*, 2010.
- [7] V. Gournay. The ductus arteriosus: physiology, regulation, and functional and congenital anomalies. *Archives of cardiovascular diseases*, 104(11):578–585, 2011.
- [8] S. A. P. Haddad and W. A. Serdijn. *Ultra low-power biomedical signal processing: an analog wavelet filter approach for pacemakers*. Springer Science & Business Media, 2009.
- [9] W. Jin-gang, S. Xiao-Bo, W. Ping, H. Wei, and D. Cui-Lian. Remote heart sound monitoring system. In *Engineering in Medicine and Biology Society, 2005. IEEE-EMBS 2005. 27th Annual International Conference of the*, pages 2138–2140. IEEE, 2006.
- [10] A. R. Kemper, W. T. Mahle, G. R. Martin, W. C. Cooley, P. Kumar, W. R. Morrow, K. Kelm, G. D. Pearson, J. Glidewell, S. D. Grosse, et al. Strategies for implementing screening for critical congenital heart disease. *Pediatrics*, 128(5):e1259–e1267, 2011.
- [11] L. I. Kuncheva. *Combining pattern classifiers: methods and algorithms*. John Wiley & Sons, 2004.
- [12] A. J. Marelli, A. S. Mackie, R. Ionescu-Ittu, E. Rahme, and L. Pilote. Congenital heart disease in the general population changing prevalence and age distribution. *Circulation*, 115(2):163–172, 2007.
- [13] M. C. White. Approach to managing children with heart disease for noncardiac surgery. *Pediatric Anesthesia*, 21(5):522–529, 2011.
- [14] S. Yuenyong, W. Kongpravechnon, K. Tungpimolrut, and A. Nishihara. *Automatic heart sound analysis for tele-cardiac auscultation*. IEEE, 2009.